

Calibration Certificate

14A0054A-CC-215-120410-1

Oxford Technical Solutions

This certificate states the performance of the product after any change to the internal IMU sensor model. This certificate may not be reproduced other than in full.

Overview

Issued by	Oxford Technical Solutions Ltd. 77 Heyford Park, Upper Heyford, Oxfordshire, OX25 5HD, UK Registered in England and Wales No. 3534778
Confidentiality	Confidential customer information

Calibration Information

Calibration by	Sample Person
Calibration ID	14A0054A-CC-215-120410
Calibration method	14A0054A
Calibration software	OxTS Calibrate, 120320.14g
Calibration date	2012-04-10
Document revision	1

Certificated Item

This calibration only applies to the product listed here:

Model	RT3102
Serial number	215
Result	Pass

IMU Calibration

The performance of the RT3102 after calibration.

Values

Measured alignment matrix and bias vector for the accelerometers and gyroscopes:

$$M_{\text{acc}} = \begin{pmatrix} +.999712 & +5.77480 \cdot 10^{-11} & -3.47950 \cdot 10^{-11} \\ -1.73477 \cdot 10^{-5} & +.999730 & -3.01018 \cdot 10^{-11} \\ +1.86725 \cdot 10^{-6} & +2.05923 \cdot 10^{-6} & +.999704 \end{pmatrix} \quad B_{\text{acc}} = \begin{pmatrix} +2.24315 \cdot 10^{-4} \\ -5.66539 \cdot 10^{-4} \\ -3.16899 \cdot 10^{-4} \end{pmatrix}$$
$$M_{\text{gyr}} = \begin{pmatrix} +1.00019 & +1.09739 \cdot 10^{-4} & -6.85871 \cdot 10^{-5} \\ -1.09739 \cdot 10^{-4} & +1.00010 & -1.37174 \cdot 10^{-4} \\ +9.60219 \cdot 10^{-5} & -2.88066 \cdot 10^{-4} & +.999945 \end{pmatrix} \quad B_{\text{gyr}} = \begin{pmatrix} +5.05873 \cdot 10^{-4} \\ -.0617306 \\ -.0345440 \end{pmatrix}$$

The matrices M_{acc} and M_{gyr} are dimensionless. The units of B_{acc} are ms^{-2} and B_{gyr} are $^{\circ}\text{s}^{-1}$.

